

SOLUTION CHANGE

DataMaster DMT: 102906
Location: Orange SO
Date: 05/09/2016
Performed by: WAYNE F BRIGGS



Accuracy and Precision Check

Concentration	= 0.099 g/210L
Lot #	= 15-13-100
Range	= 0.094 - 0.104
Average	= 0.097 g/210L
Std Dev	= 0.0005

*RCD
5/9/16*

Simulator Temperature: 33.9°C

Performed by WFB

Date 05/09/2016