

# SOLUTION CHANGE

DataMaster DMT: 103509  
Location: Hardwick-Greensboro PD  
Date: 10/15/2015  
Performed by: CHRISTOPHER R TETREAULT



## Accuracy and Precision Check

Concentration = 0.100 g/210L  
Lot # = 15-07-100  
Range = 0.095 - 0.105  
Average = 0.097 g/210L  
Std Dev = 0.0005

Simulator Temperature: 34.1°C

Performed by CR Tetreault  
ROD 10/15/15

Date 10/15/2015