

SOLUTION CHANGE

DataMaster DMT: 103509
Location: Hardwick-Greensboro PD
Date: 12/18/2015
Performed by: CHRISTOPHER R TETREAULT



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-13-100
Range = 0.094 - 0.104
Average = 0.096 g/210L
Std Dev = 0.0008

ROD
12/21/15

Simulator Temperature: 34.2°C

Performed by

CR Tetreault

Date

12/18/2015