

SOLUTION CHANGE

DataMaster DMT: 103509
Location: Hardwick-Greensboro PD
Date: 04/14/2016
Performed by: CHRISTOPHER R TETREAULT



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-13-100
Range = 0.094 - 0.104
Average = 0.098 g/210L
Std Dev = 0.0004

RCD 4/14/16

Simulator Temperature: 34.1°C

Performed by *CR*

Date 04/14/2016