

# SOLUTION CHANGE

DataMaster DMT: 103509  
Location: Hardwick-Greensboro PD  
Date: 09/17/2016  
Performed by: STEVEN MITCHELL



## Accuracy and Precision Check

Concentration = 0.098 g/210L  
Lot # = 16-04-100  
Range = 0.093 - 0.103  
Average = 0.098 g/210L  
Std Dev = 0.0004

*RCD 9/19/16*

Simulator Temperature: 34.2°C

Performed by

*Steven Mitchell*

Date

09/17/2016