Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
04/24/2015	23:04:37	ST. JOHNSBURY PD	ADRIAN L HAHR	0.084	0.077	0.095	0.095	
04/25/2015	23:43:12	ST. JOHNSBURY PD	ADRIAN L HAHR			0.093		Simulator Out of Range
04/25/2015	23:46:31	ST. JOHNSBURY PD	ADRIAN L HAHR	REFUSED		0.094		
04/30/2015	23:38:59	ST. JOHNSBURY PD	ADRIAN L HAHR			0.091		Simulator Out of Range

DATAMASTER DMT MAINTENANCE LOG

ST. Johnsbury Police Dept
Department/Agency

103609
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/7/14	[For RPC or Simulator Solution Change Place Label Here] Reset Clock on DMT due to time change	DMT Supervisor Sgt SJPD [signature]
12/7/14	Performed Accuracy & Precision Test "Found out of Range"	DMT Supervisor Sgt SJPD [signature]
12/7/14	[For RPC or Simulator Solution Change Place Label Here] **DataMaster Simulator Solution** Lot # 14-05-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 6/19/2014 Expires: 6/19/2015 OPENED: 12-6-14 [signature] Solution Change	DMT 12/7/14 Supervisor [signature] Sgt SJPD
02/20/15	[For **DataMaster Simulator Solution** Place Label Here] Lot # 14-14-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/16/2014 Expires: 10/16/2015 OPENED: [signature] 02-20-15 RPC	DMT Supervisor Sgt [signature] RPC 2/20/15 SJPD
5/5/15	[For **DataMaster Simulator Solution** Place Label Here] Lot # 14-14-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/16/2014 Expires: 10/16/2015 OPENED: [signature] 5/5/15 Solution Change .101	SGT. Lester Cleary [signature] DMT SUPERVISOR

ALC 803 REV 03 08/2011

DATAMASTER DMT OPERATOR USE LOG

St. Johnsbury P.D
Department/Agency

103609
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
12/8/14	Adrian Hahr	Matthew Gibson	SOR	
12/7/14	Simulator Solution Change	Mark Bickford/DMT	.100	
12/8/14	Johnson	Seth Stoney	.094	
1-1-15	Buck	Nicholas Boivin	.098	
01-31-15	Johnson	Dalon Phillips	.096	
02/06/15	Johnson	Rachel Demers ~~Eric Morrison~~	.096	
02/07/1	Johnson	Eric Morrison	.098	
02-20-15	Mark Bickford	Test/RPC	.099	
03/02/15	Haley	Mark Whitehead	.099	.099
3-17-15	Hazard	Sean Dunner	.099	
3/20/15	Upson	William Cowron	.095	.095
3/25/15	Bickford	Ethan Sawyer	.095	
3/25/15	Bickford	Morgan Lane	.096	
3/26/15	Herbes	Krystal Whitcomb	.098	.097
4/4/15	Molleur	John Poiest	.097	.096
4/18/15	Hahr	Christopher Nerhale	.096	.096
04-24-15	Hahr	Rodney Sayles	.095	.095
04-25-15	Hahr	Ryan Milligan	SOR	.094 Refusal
04-30-15	Hahr	Mark Maskell Jr	SOR	
5/3/15	Johnson	Belinda Plymak	SOR	

ALC 603 Rev 3 08/11

DATAMASTER DMT OPERATOR USE LOG

St Johnsbury PD
Department/Agency

103609
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
5/3/15	Skelton	Jayde Smith	SOR	
5/5/15	Cleary	Solution Change	.101	

ALC 603 Rev 3 08/11