Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
06/05/2015	23:25:35	LYNDONVILLE PD	ROBERT L TRUCOTT	0.133		0.094	0.093	Simulator Out of Range
06/20/2015	23:33:25	ST. JOHNSBURY PD	ADRIAN HAHR			0.093		Simulator Out of Range
06/20/2015	23:36:34	ST. JOHNSBURY PD	ADRIAN HAHR			0.094		RESPONSE TIMEOUT
06/20/2015	23:41:27	ST. JOHNSBURY PD	ADRIAN HAHR	0.162	0.164	0.094	0.094	
06/26/2015	09:47:35	SJPD	MARK R BICKFORD	0.000		0.099		

Status Record History

Status Date	Status Time	Test Type	Status Message
06/05/2015	23:31:17	DUI Subject Record	Simulator Out of Range
06/20/2015	23:34:55	DUI Subject Record	Simulator Out of Range
06/20/2015	23:39:56	DUI Subject Record	RESPONSE TIMEOUT

DATAMASTER DMT MAINTENANCE LOG

ST Johnsbury P.D.
Department/Agency

103609
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/6/15	Performed Accuracy & Precision Test on DMT [For RPC or Simulator Solution Change Place Label Here] "Simulator out of Range"	DMT Supervisor Sgt/SJPD
6/26/15	[For RPC or Simulator Solution Change Place Label Here] **DataMaster Simulator Solution** Lot # 15-06-100 EtOHaq Conc: 0.100 g/210L Range: 0.095 - 0.105 Preparation Date: 4/28/2015 Expires: 4/28/2016 OPENED: 6-26-15	DMT Supervisor Sgt/SJPD RPC
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011

DATAMASTER DMT OPERATOR USE LOG

St Johnsbury PD
Department/Agency

103609
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
5/3/15	Skelton	Jayde Smith	SOR	
5/5/15	Cleary	Solution Change	.101	
5/10/15	Loomis	Keith Nickles	.099	
5/17/15	UPSON	Lucas Barillard	.098	
5/22/15	Rivard	Samantha Peuolo	.099	
6-5-15	Tirrell	Edwin Chamid	.099	SOR
06-20-15	Hahn	Brendan Kozlowski	SOR	Time Out
06-20-15	Hahn	Brendan Kozlowski	.094	.094
06-26-15	Mark Bickford	Test / RPC	.100	
06-26-15	Mark Bickford	Mark Bickford	.100	
07-04-15	Johnson	Jennifer Schafer	.098	
07-4-15	Warren	Kimbley Hastings	.098	

ALC 603 Rev 3 08/11

ROUTINE PERFORMANCE CHECK REPORT

DataMaster DMT: 103609
Location: St. Johnsbury PD
Calibration Date: 09/21/2011
Certification Date: 09/21/2011
Installation Date: 01/09/2013
RPC Date: 06/26/2015
Supervisor Name: MARK R BICKFORD

Diagnostic Results

```
VERSIONS
DMT: 1.01
PIC: 2.06
Modem: 2.4
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.8°C
Breath Tube    = 48.1°C
Digital Sim    = 34.1°C

SETTINGS
Lamp Voltage   = 1.68 V
Cooler Voltage = 1.66 V
Bias Voltage   = 81 V
Chopper Freq   = 538 Hz

PUMP INFO
Flow Rate = 5.514 L/M

DETECTOR INFO
PUMP     ON        OFF
MAX(V)   0.5831    0.5864
MIN(V)   0.5821    0.5850

FILTER INFO
Filter 1   0.585  Zero = true
Filter 2   0.796  Zero = true
Filter 3   0.497  Zero = true

CALIBRATION CHECK
Xq  = 0.083  1.17%
```

Routine Performance Check Passed

Accuracy and Precision Check

Concentration	= 0.100 g/210L
Lot #	= 15-06-100
Range	= 0.095 - 0.105
Average	= 0.100 g/210L
Std Dev	= 0.0000

RF Detection Test

Passed

Sample Acceptance Test

Passed



Performed by Mark Bickford Date 06/26/2015

Reviewed by _____ Date _____



CHECK-IN CONFIRMATION TEST

STATE OF VERMONT

DataMater DMT: 103609
Location: St. Johnsbury PD
Certification Date: 09/21/2011
Installation Date: 01/09/2013

Date: 06/26/2015
Time: 09:47:35

SUBJECT NAME:
MARK R BICKFORD
DOB: 10/08/1967 AGE: 47
GENDER: M
LICENSE #: 80958005 VT

TEST OPERATOR NAME:
MARK R BICKFORD
AGENCY: SJPD

BLANK TEST	0.000	09:48
CALIBRATION CHECK	PASSED	09:48
SIMULATOR VAPOR 34.0°C	0.099	09:48
SUBJECT SAMPLE (Vol=2.9L)	0.000	09:49
BLANK TEST	0.000	09:50

All alcohol values reported in g/210L
Sim. Vapor Concentration 0.100
Sim. Vapor Range 0.095 - 0.105
Lot #: 15-06-100

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 103609
Location: St. Johnsbury PD

Date: 06/06/2015
Time: 08:02:47

SUPERVISOR NAME:
MARK R BICKFORD

SOLUTION LOT #: 14-14-100
SOLUTION CONCENTRATION: 0.099

BLANK TEST	0.000	08:03
CALIBRATION CHECK	PASSED	08:03
SIMULATOR VAPOR 34.0°C	0.096	08:04
SIMULATOR VAPOR 34.0°C	0.096	08:05
SIMULATOR VAPOR 34.0°C	0.097	08:06
SIMULATOR VAPOR 34.0°C	0.097	08:07
SIMULATOR VAPOR 34.0°C	0.097	08:08
SIMULATOR VAPOR 34.0°C	0.097	08:09
SIMULATOR VAPOR 33.9°C	0.097	08:10
SIMULATOR VAPOR 34.1°C	0.097	08:11
SIMULATOR VAPOR 34.0°C	0.097	08:13
SIMULATOR VAPOR 34.0°C	0.097	08:14
BLANK TEST	0.000	08:14

Average = 0.096
Std Dev = 0.000





DUI SUBJECT TEST

STATE OF VERMONT

DataMater DMT: 103609
Location: St. Johnsbury PD
Certification Date: 09/21/2011
Installation Date: 01/09/2013

Date: 06/05/2015
Time: 23:25:35

SUBJECT NAME:
[redacted]
DOB: [redacted] AGE: [redacted]
GENDER: [redacted]
LICENSE #: [redacted]

CASE #: 15LV00464
TIME OF OPERATION: 22:03
LOCATION OF INCIDENT:
LILY POND ROAD
TOWN: LYNDON
COUNTY: CALEDONIA

TEST OPERATOR NAME:
ROBERT L TRUCOTT
AGENCY: LYNDONVILLE PD

Obs. Period Start:		23:10
BLANK TEST	0.000	23:26
CALIBRATION CHECK	PASSED	23:26
SIMULATOR VAPOR 34.0°C	0.094	23:26
BLANK TEST	0.000	23:27
SUBJECT SAMPLE (Vol=4.4L)	0.133	23:28
BLANK TEST	0.000	23:30
CALIBRATION CHECK	PASSED	23:30
SIMULATOR VAPOR 34.0°C	0.093	23:31

Simulator Out of Range

All alcohol values reported in g/210L
Sim. Vapor Concentration 0.099
Sim. Vapor Range 0.094 - 0.104
Lot #: 14-14-100

DUI SUBJECT TEST

STATE OF VERMONT

DataMater DMT: 103609
Location: St. Johnsbury PD
Certification Date: 09/21/2011
Installation Date: 01/09/2013

Date: 06/20/2015
Time: 23:33:25

SUBJECT NAME:
[redacted]
DOB: [redacted] AGE: [redacted]
GENDER: [redacted]
LICENSE #: [redacted]

CASE #: 15SJ003879
TIME OF OPERATION: 22:46
LOCATION OF INCIDENT:
ROUTE 5
TOWN: ST. JOHNSBURY
COUNTY: CALEDONIA

TEST OPERATOR NAME:
ADRIAN HAHR
AGENCY: ST. JOHNSBURY PD

Obs. Period Start:		23:18
BLANK TEST	0.000	23:34
CALIBRATION CHECK	PASSED	23:34
SIMULATOR VAPOR 34.0°C	0.093	23:34

Simulator Out of Range

All alcohol values reported in g/210L
Sim. Vapor Concentration 0.099
Sim. Vapor Range 0.094 - 0.104
Lot #: 14-14-100

DUI SUBJECT TEST

STATE OF VERMONT

DataMater DMT: 103609
Location: St. Johnsbury PD
Certification Date: 09/21/2011
Installation Date: 01/09/2013

Date: 06/20/2015
Time: 23:36:34

SUBJECT NAME:
[redacted]
DOB: [redacted] AGE: [redacted]
GENDER: [redacted]
LICENSE #: [redacted]

CASE #: 15SJ003879
TIME OF OPERATION: 22:46
LOCATION OF INCIDENT:
ROUTE 5
TOWN: ST. JOHNSBURY
COUNTY: CALEDONIA

TEST OPERATOR NAME:
ADRIAN HAHR
AGENCY: ST. JOHNSBURY PD

Obs. Period Start:		23:18
BLANK TEST	0.000	23:37
CALIBRATION CHECK	PASSED	23:37
SIMULATOR VAPOR 33.9°C	0.094	23:38
BLANK TEST	0.000	23:38

RESPONSE TIMEOUT

All alcohol values reported in g/210L
Sim. Vapor Concentration 0.099
Sim. Vapor Range 0.094 - 0.104
Lot #: 14-14-100

DUI SUBJECT TEST

STATE OF VERMONT

DataMater DMT: 103609
Location: St. Johnsbury PD
Certification Date: 09/21/2011
Installation Date: 01/09/2013

Date: 06/20/2015
Time: 23:41:27

SUBJECT NAME:
[redacted]
DOB: [redacted] AGE: [redacted]
GENDER: [redacted]
LICENSE #: [redacted]

CASE #: 15SJ003879
TIME OF OPERATION: 22:46
LOCATION OF INCIDENT:
ROUTE 5
TOWN: ST. JOHNSBURY
COUNTY: CALEDONIA

TEST OPERATOR NAME:
ADRIAN HAHR
AGENCY: ST. JOHNSBURY PD

Obs. Period Start:		23:18
BLANK TEST	0.000	23:42
CALIBRATION CHECK	PASSED	23:42
SIMULATOR VAPOR 34.0°C	0.094	23:42
BLANK TEST	0.000	23:43
SUBJECT SAMPLE (Vol=4.0L)	0.162	23:44
BLANK TEST	0.000	23:45
CALIBRATION CHECK	PASSED	23:45
SIMULATOR VAPOR 34.0°C	0.094	23:46
BLANK TEST	0.000	23:47
SUBJECT SAMPLE (Vol=4.6L)	0.164	23:47

All alcohol values reported in g/210L
Sim. Vapor Concentration 0.099
Sim. Vapor Range 0.094 - 0.104
Lot #: 14-14-100