INSTALLATION REPORT

DataMaster DMT: 103709
Location: St. Johnsbury VSP
Calibration Date: 09/08/2011
Certification Date: 09/08/2011
Installation Date: 09/30/2011
Installed by: STEVEN E HARNOIS

Site meets specification in doc P-ALC-210

Default Options Loaded

Diagnostic Results

```
VERSIONS
DMT: 1.00
PIC: 2.06
Modem: 2.1
Questions: 2.0

TEMPERATURES

Sample Chamber = 48.7°C
Breath Tube    = 48.1°C
Digital Sim    = 33.9°C

SETTINGS
Lamp Voltage   = 1.75 V
Cooler Voltage = 1.71 V
Bias Voltage   = 80 V
Chopper Freq   = 534 Hz

PUMP INFO
Flow Rate = 6.043 L/M

DETECTOR INFO
PUMP     ON        OFF
MAX(V)   0.0484    0.0512
MIN(V)   0.0467    0.0502

FILTER INFO
Filter 1   0.051  Zero = true
Filter 2   0.159  Zero = true
Filter 3   0.419  Zero = true

CALIBRATION CHECK
Xq  = 0.075  0.23%
```

Accuracy and Precision Check

Concentration	= 0.102 g/210L
Lot #	= 12-13-100
Range	= 0.097 - 0.107
Average	= 0.103 g/210L
Std Dev	= 0.0000

RF Detection Test

Passed

Sample Acceptance Test

Passed



INSTALLATION PASSED

Performed by _____ Date 09/30/2011 10-4-11

Reviewed by _____ Date 09/30/11

INSTALLATION REPORT

DataMaster DMT: 103709
Location: St. Johnsbury VSP
Calibration Date: 09/08/2011
Certification Date: 09/08/2011
Installation Date: 09/30/2011
Installed by: STEVEN E HARNOIS

Site meets specification in doc P-ALC-210

Default Options Loaded

Diagnostic Results

```
VERSIONS
DMT: 1.00
PIC: 2.06
Modem: 2.1
Questions: 2.0

TEMPERATURES

Sample Chamber = 48.9°C
Breath Tube    = 48.1°C
Digital Sim    = 33.9°C

SETTINGS
Lamp Voltage   = 1.75 V
Cooler Voltage = 1.71 V
Bias Voltage   = 80 V
Chopper Freq   = 533 Hz

PUMP INFO
Flow Rate = 5.916 L/M

DETECTOR INFO
PUMP     ON        OFF
MAX(V)   0.0445    0.0472
MIN(V)   0.0431    0.0461

FILTER INFO
Filter 1   0.046  Zero = true
Filter 2   0.154  Zero = true
Filter 3   0.413  Zero = true

CALIBRATION CHECK
Xq  = 0.075  0.12%
```

Accuracy and Precision Check

Concentration =
Lot # =
Range =
Average =
Std Dev =

Simulator Timed Out. 09/30/11 SEH

9/30/11 SEH

10-4-11

INSTALLATION FAILED

Performed by _____ Date 09/30/2011

Reviewed by _____ Date 09/30/11