



VERMONT FORENSIC LABORATORY

PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
<http://vfl.vermont.gov>

DATAMASTER TECHNICAL SUPPORT INQUIRY

Agency/Site: St. Johnsbury VSP DataMaster Serial #: 103709

Date: 1/19/2013 Requestor: Seth Loomis

Instrument Supervisor: Seth Loomis Phone #: (802) 748-3111

Taken By: R. Driscoll

Problem: Simulator jar chipped and does not seal properly _____

Laboratory Response: R. Driscoll will investigate on site _____

Work Performed: A large piece of the simulator jar had broken. Replaced jar and pressure tested for leaks. Simulator head was leaking so replaced simulator head as well (s/n: G4861). Assigned simulator to DMT following the procedure outlined in P-ALC 203. Temperatures matched using NIST traceable thermometer s/n 111814522. DataMaster left out of service for DMT supervisor to perform their February RPC. _____

On Site In-house Attachments: 1 page

Performed By: R. Driscoll RCD Date: 1/22/2013
Final Review: [Signature] Date: 1/24/2013

Driscoll, Rob

From: Loomis, Seth
Sent: Saturday, January 19, 2013 6:37 PM
To: Driscoll, Rob
Cc: Herbes, Cassandra
Subject: DMT Solution Jar

Mr. Driscoll,

Last night while changing out the solution for the DMT, it was observed there was a chip in the top of the jar. The DMT has been taken out of service for now, however, we will need a replacement jar as there is not a solid seal. If there is someone else I need to contact regarding this issue please let me or Trooper Herbes know. Below I provided the barracks mailing address. Thank you for your time.

Respectfully,

Trooper Seth Loomis
Vermont State Police
"B" Troop- St. Johnsbury

1068 U.S. Route 5; Ste. 1
St. Johnsbury, VT 05819

(802) 748-3111

Cell: (518) 322-6334

Email: seth.loomis@state.vt.us