ROUTINE PERFORMANCE CHECK REPORT

DataMaster DMT: 103709
Location: St. Johnsbury VSP
Calibration Date: 09/08/2011
Certification Date: 09/08/2011
Installation Date: 09/30/2011
RPC Date: 10/29/2014
Supervisor Name: CASSANDRA L HERBES

Diagnostic Results

```
VERSIONS
DMT: 1.01
PIC: 2.06
Modem: 2.4
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.7°C
Breath Tube    = 48.1°C
Digital Sim    = 33.9°C

SETTINGS
Lamp Voltage   = 1.75 V
Cooler Voltage = 1.71 V
Bias Voltage   = 80 V
Chopper Freq   = 545 Hz

PUMP INFO
Flow Rate = 6.068 L/M

DETECTOR INFO
PUMP     ON        OFF
MAX(V)   0.5896    0.5915
MIN(V)   0.5882    0.5900

FILTER INFO
Filter 1   0.590   Zero = true
Filter 2   0.689   Zero = true
Filter 3   0.935   Zero = true

CALIBRATION CHECK
Xq  = 0.076  1.73%
```

Routine Performance Check Passed

Accuracy and Precision Check

Concentration	= 0.098 g/210L
Lot #	= 14-05-100
Range	= 0.093 - 0.103
Average	= 0.101 g/210L
Std Dev	= 0.0000

RF Detection Test

Passed

Sample Acceptance Test

Passed



Performed by _____ Date 10/29/2014

Reviewed by _____ Date 11/3/14

11/3/2014