



**VERMONT FORENSIC LABORATORY**

PO BOX 47  
WATERBURY, VERMONT 05676-0047  
TEL: 802-244-8788  
FAX: 802-241-5557  
<http://vfl.vermont.gov>

**DATAMASTER TECHNICAL SUPPORT INQUIRY**

<b>Agency/Site:</b> <u>Windsor PD</u>	<b>DataMaster Serial #:</b> <u>104309</u>
Date: <u>2/20/2013</u>	Requestor: <u>Jonathan Adams</u>
Instrument Supervisor: <u>Jonathan Adams</u>	Phone #: <u>(802) 674-9042</u>
Taken By: <u>R. Driscoll</u>	
Problem: Simulator solution jar broken during RPC. _____ _____ _____	
Laboratory Response: Will bring replacement simulator jar to agency. _____ _____	
Work Performed: R. Driscoll replaced simulator jar and leak checked. Simulator leaking so replaced with spare unit. After solution temperature stabilized, the temp was matched to the DMT using NIST thermometer S/N 111814522. Performed RPC protocol as it was due. _____ _____ _____ _____ _____ _____ _____ _____	
<input checked="" type="checkbox"/> On Site <input type="checkbox"/> In-house	Attachments: <u>0</u> pages
Performed By: <u>R. Driscoll</u> <i>RD</i>	Date: <u>2/20/2013</u>
Final Review: <u><i>[Signature]</i></u>	Date: <u>2/20/2013</u>

*Page 1 of 1 RD*