

# SOLUTION CHANGE

DataMaster DMT: 104409  
Location: Winooski PD  
Date: 09/08/2015  
Performed by: ROBERT DRISCOLL



## Accuracy and Precision Check

Concentration = 0.099 g/210L  
Lot # = 15-09-100  
Range = 0.094 - 0.104  
Average = 0.099 g/210L  
Std Dev = 0.0015

Simulator Temperature: 34.0°C

Performed by

*RD*

Date

09/08/2015

# SOLUTION CHANGE

DataMaster DMT: 104409  
Location: Winooski PD  
Date: 09/08/2015  
Performed by: ROBERT DRISCOLL



## Accuracy and Precision Check

Concentration  
Lot #  
Range  
Average  
Std Dev

Simulator Temperature:

Solution Change Failed

*Simulator timed out 9/8/15  
ROD*

Performed by

Handwritten signature of Robert Driscoll.

Date

09/08/2015