SOLUTION CHANGE

DataMaster DMT: 104609
Location: UVM PD
Date: 12/22/2015
Performed by: MATTHEW L COLLINS

Accuracy and Precision Check

Concentration	= 0.097 g/210L
Lot #	= 15-23-100
Range	= 0.092 - 0.102
Average	= 0.098 g/210L
Std Dev	= 0.0004

RCD 12/23/15

Simulator Temperature: 34.1°C

Performed by _____ Date 12/22/2015