SOLUTION CHANGE

DataMaster DMT: 104809
Location: St. Albans PD
Date: 12/22/2015
Performed by: MICHAEL C SCHRADER

Accuracy and Precision Check

Concentration	= 0.100 g/210L
Lot #	= 15-07-100
Range	= 0.095 - 0.105
Average	= 0.101 g/210L
Std Dev	= 0.0004

RCD 12/22/15

Simulator Temperature: 34.0°C

Performed by M. Schrader Date 12/22/2015

SOLUTION CHANGE

DataMaster DMT: 104809
Location: St. Albans PD
Date: 12/22/2015
Performed by: MICHAEL C SCHRADER

Accuracy and Precision Check

Concentration	= 0.100 g/210L
Lot #	= 15-07-100
Range	= 0.095 - 0.105
Average	= 0.091 g/210L
Std Dev	= 0.0058

Standard Deviation out of range

Simulator Temperature: 34.0°C

Solution Change Failed

Simulator jar was not sealing properly during first solution change attempt. DMT Supervisor was able to tighten the jar and fix the leak. RCD 12/22/15

Performed by M. Schrader **Date** 12/22/2015

2 of 2 RCD