

SOLUTION CHANGE

DataMaster DMT: 104809
Location: St. Albans PD
Date: 03/30/2016
Performed by: MICHAEL C SCHRADER



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-24-100
Range = 0.094 - 0.104
Average = 0.102 g/210L
Std Dev = 0.0007

*RCP
3/30/16*

Simulator Temperature: 33.9°C

Performed by

M. C. Schrader
104809

Date

03/30/2016