

# SOLUTION CHANGE

DataMaster DMT: 104809  
Location: St. Albans PD  
Date: 05/24/2016  
Performed by: MICHAEL C SCHRADER



## Accuracy and Precision Check

Concentration = 0.099 g/210L  
Lot # = 15-24-100  
Range = 0.094 - 0.104  
Average = 0.102 g/210L  
Std Dev = 0.0005

*RCD  
5/24/16*

Simulator Temperature: 34.0°C

Performed by

*M. Schrader*  
PH 17-155

Date

05/24/2016