# SOLUTION CHANGE

DataMaster DMT: 104909
Location: Williston PD
Date: 11/27/2015
Performed by: ERIC D SHEPARD

**Accuracy and Precision Check**

| | |
|---|---|
| Concentration | = 0.100 g/210L |
| Lot # | = 15-07-100 |
| Range | = 0.095 - 0.105 |
| Average | = 0.099 g/210L |
| Std Dev | = 0.0004 |

KCD 11/30/15

Simulator Temperature: 34.1°C

Performed by _____ Date 11/27/2015