

SOLUTION CHANGE

DataMaster DMT: 121506
Location: Swanton PD
Date: 02/12/2016
Performed by: ANDREW G UNDERWOOD



Accuracy and Precision Check

Concentration = 0.100 g/210L
Lot # = 15-25-100
Range = 0.095 - 0.105
Average = 0.098 g/210L
Std Dev = 0.0004

*RCD
2/12/16*

Simulator Temperature: 33.9°C

Performed by *Andrew Underwood*

Date 02/12/2016