

VERMONT FORENSIC LABORATORY

PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
<http://vfl.vermont.gov>



DataMaster DMT Site Evaluation Checklist

Agency or Site Information

Name: Vergennes 2D

Address: 8 Main Street
Vergennes, Vermont 05491

Phone: 802-877-1152

Fax: 802-877-1155

Personnel Information

Head of Site: George F. Merke /

Current DataMaster Supervisor (s):
Bred Newson

Proposed DataMaster Supervisor(s):

(Proposed) DataMaster DMT Environment

Limited Controlled Access: Y N

Low Activity Area: Y N

Unlimited Access by Other Law Enforcement Personnel: Y N

Access by VFL Technical Services: Y N

Air Conditioned: Y N

Room Temperature 65° - 85° F: Y N

Adequate Ventilation: Y N

Free of Cleaning Supplies: Y N

Clean: Y N

(Proposed) DataMaster DMT Connection(s)

Power:
Min 15 Amp Y N

Limited low power sharing: Y N

Ethernet Available: Y N

Distance to Ethernet: < 1'

(Proposed) DataMaster DMT Site Maintenance Agreement

Will notify VFL Alcohol Program staff of maintenance and supply needs promptly: Y N

Will notify VFL Alcohol Program staff of changes in building conditions or moves promptly: Y N

Will notify VFL Alcohol Program staff of relevant personnel changes at site promptly: Y N

VFL Representative: Robert Driscoll [Signature] 3/5/14
Print Name Signature Date

Agency Representative: George F. Merke [Signature] 03/05/14
Print Name Signature Date