VERMONT FORENSIC LABORATORY

PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
http://vfl.vermont.gov

DATAMASTER TECHNICAL SUPPORT INQUIRY

Agency/Site: Vergennes PD **DataMaster Serial #:** 122106

Date: 1/21/2015 Requestor: Brent Newton

Instrument Supervisor: Brent Newton Phone #: (802)877-2201

Taken By: Robert Driscoll

Problem: Simulator solution out of range post APM.

Laboratory Response: R. Driscoll will inspect on site.

Work Performed: Simulator was slightly leaking during APM. Replaced simulator with Guth S/N: G11306 and matched temperatures using NIST traceable thermometer S/N: 111814517. Lastly, the February RPC protocol was performed to complete the solution change.

☒ On Site ☐ In-house Attachments: 0 pages

Performed By: Robert Driscoll RCD Date: 1/22/2015

Final Review: [signature] Date: 1/22/2015

Page 1 of 1 RCD