DATAMASTER DMT MAINTENANCE LOG

Jan 2015

FCSD - Enosburg
Department/Agency

1222-06
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/12/14	**DataMaster Simulator Solution** [For RPC ce Label Here] Lot # 14-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 5/27/2014 Expires: 5/27/2015 OPENED: 10/12/14 RPC	Chb
1/9/15	**DataMaster Simulator Solution** [For RPC Label Here] Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: RCD 1/9/15 out of range during AOM, performed solution change + rerun AOM protocol	R. Driscoll DMT TECH.
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011

DATAMASTER DMT OPERATOR USE LOG

FCSO Enosburg
Department/Agency

122206
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
7/18/14	Miles /C	Abby Maybee	.099	
7/16/14	Miles /C	Brian Barber		
8-8-14	Chiavarelle	Colin Sadler	.098	
8-15-14	Dinitz	Robert Bushey	.099	
8/18/14	Vaillancourt	Dennis Kent	.099	.099
8/20/14	D. Scott	Mary Campbell	.099	
9/2/14	Miles, C	Michael Gould	.097	
9/10/14	Griswold	Eric Jette	.099	
10/5/14	Dinitz	Clements C		
10/12/14	Miles /C	RPC	.099	
10/19/14	Vaillancourt	Evelyn Therrien	.101	.101
11/11/14	Griswold	Jessica Taylor	.099	.099
12/25/14	Vaillancourt	Walter Wagner	.093 (out of range)	
1/9/15	R. Driscoll	TEST/APM	.093 out of range	
1/9/15	R. Driscoll	TEST /solution change	.099	—
1/9/15	R. Driscoll	TEST / APM	.099	—
1/11/15	M. Henry	Nicholas Gerrow	.097	—
1/12/15	Miles /C	Luke Danyow	.097	.096

ALC 603 Rev 3 08/11

Sup 1st CM

Breath Record History

FCSD - Enosburg
Jan 2015 122206

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
11/11/2014	23:58:04	FRANKLIN COUNTY SO	COREY L GRISWOLD	0.064	0.063	0.097	0.098	
12/25/2014	20:46:42	FRANKLIN COUNTY SO	NATHAN VAILLANCOURT			0.093		Simulator Out of Range

Status Record History

FCSD - 122206
Jan 2015 100135

Status Date	Status Time	Test Type	Status Message
12/25/2014	20:48:11	DUI Subject Record	Simulator Out of Range

Log On Record History

FCSD - Enosburg
Jan 2015 122206

Test Date	Test Time	Name