Datamaster Reports

11/01/14 – 03/31/15

FCSO Enosburg Outpost

Device # 122206

RECEIVED
FRANKLIN COUNTY SAO
APR 07 2015

Log On Record History

FCSD - Enosburg
11/1/14 - 3/31/15
122206

Test Date	Test Time	Name

Status Record History

FCSD - Enosburg
11/1/14 - 3/31/15
122206

Status Date	Status Time	Test Type	Status Message
12/25/2014	20:48:11	DUI Subject Record	Simulator Out of Range

Breath Record History

FCSD - Enosburg
11/1/15 - 3/31/15
122206

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
11/11/2014	23:58:04	FRANKLIN COUNTY SO	COREY L GRISWOLD	0.064	0.063	0.097	0.098	
12/25/2014	20:46:42	FRANKLIN COUNTY SO	NATHAN VAILLANCOURT			0.093		Simulator Out of Range

DATAMASTER DMT MAINTENANCE LOG

11/1/14 - 3/31/15

FCSD - Enosburg
Department/Agency

122206
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/12/14	**DataMaster Simulator Solution** [For RPC ce Label Here] Lot # 14-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 5/27/2014 Expires: 5/27/2015 OPENED: 10/12/14 RPC	Chb
1/9/15	**DataMaster Simulator Solution** [For RPC Label Here] Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: RCD 1/9/15 out of range during APM, performed solution change + rerun APM protocol.	R. Driscoll DMT TECH.
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011

DATAMASTER DMT OPERATOR USE LOG

FCSO Enosburg — 11/1/14 - 3/31/15 — 122206

Department/Agency — DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
7/18/14	Miles /C	Abby Mayhew	.099	
7/16/14	Miles /C	Brian Barber		
8-8-14	Chiaravalle	Colin Sadler	.098	
8-15-14	Diniz	Robert Bushey	.099	
8/18/14	Vaillancourt	Dennis Kent	.099	.099
8/20/14	D. Scott	Mary Campbell	.099	
9/2/14	Miles, C	Michael Gould	.097	
9/10/14	Griswold	Eric Jette	.099	
10/5/14	Diniz	Clements C		
10/12/14	Miles /C	RPC	.099	
10/19/14	Vaillancourt	Evelyn Therrien	.101	.101
11/11/14	Griswold	Jessica Taylor	.099	.099
12/25/14	Vaillancourt	Walter Wagner	.093 (out of range)	
1/9/15	R. Driscoll	TEST / APM	.093 out of range	—
1/9/15	R. Driscoll	TEST / solution change	.099	—
1/9/15	R. Driscoll	TEST / APM	.099	—
1/11/15	M. Henry	Nicholas Gerrow	.097	—
1/12/15	Miles /C	Luke Danyow	.097	.096

ALC 603 Rev 3 08/11

Sup 1st cm

Breath Record History

FCSD - Enosburg
11/1/14 - 3/31/15
122206

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
01/11/2015	02:45:12	ST. JOHNSBURY VSP	MICAHEL J HENRY	0.156		0.097		
01/12/2015	23:39:12	FRANKLIN COUNTY SO	CHAD A MILES	0.130	0.133	0.097	0.096	
03/13/2015	21:35:09	FRANKLIN COUNTY SO	NATHAN VAILLANCOURT	0.092		0.098		
03/14/2015	05:03:31	ST. ALBANS VSP	BENJAMIN M PATNODE	0.117		0.096		
03/29/2015	18:20:12	FRANKLIN COUNTY SO	NATHAN VAILLANCOURT	0.142	0.152	0.096	0.096	

Status Record History

FCSD - Enosburg
11/1/14 - 3/31/15
122206

Status Date	Status Time	Test Type	Status Message
01/09/2015	09:01:38	APM Record	ACCURACY AND PRECISION CHECK FAILED

Log On Record History

FCSD - Enosburg
11/1/14 - 3/31/15
122206

Test Date	Test Time	Name

DATAMASTER DMT OPERATOR USE LOG

FCSO Enosburg
Department/Agency

11/1/14 - 3/31/15

122206
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
7/18/14	Miles /C	Abby Mayhew	.099	
7/16/14	Miles /C	Brian Barber		
8-8-14	Chiaravalle	Colin Sadler	.098	
8-15-14	Diniz	Robert Bushey	.099	
8/18/14	Vaillancourt	Dennis Kent	.099	.099
8/22/14	D. Scott	Mary Campbell	.099	
9/2/14	Miles, C	Michael Gould	.097	
9/10/14	Griswold	Eric Jette	.099	
10/5/14	Diniz	Clements C		
10/12/14	Miles /C	RPC	.099	
10/19/14	Vaillancourt	Evelyn Therrien	.101	.101
11/11/14	Griswold	Jessica Taylor	.099	.099
12/25/14	Vaillancourt	Walter Wagner	.093 (out of range)	
1/9/15	R. Driscoll	TEST / APM	.093 out of range	
1/9/15	R. Driscoll	TEST / solution change	.099	—
1/9/15	R. Driscoll	TEST / APM	.099	—
1/11/15	M. Henry	Nicholas Gerrow	.097	—
1/12/15	Miles /C	Luke Banyow	.097	.096
02/15/15	N. Vaillancourt	Test / RPC	.099	
03/13/15	N. Vaillancourt	Travis Barbour	.098	

ALC 603 Rev 3 08/11

Sup 1st 8m

DATAMASTER DMT OPERATOR USE LOG

FCSO Enosburg — Department/Agency

11/1/14 - 3/31/15

122206 — DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
3/14	B. PATNODE	DEVON GOSS	.096	
03/29	N. Vaillancourt	Erik Hance	.096	.096

DATAMASTER DMT MAINTENANCE LOG

11/1/14 - 3/31/15

FCSD - Enosburg
Department/Agency

122206
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/12/14	**DataMaster Simulator Solution** [For RPC ce Label Here] Lot # 14-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 5/27/2014 Expires: 5/27/2015 OPENED: 10/12/14 RPC	Chb
1/9/15	**DataMaster Simulator Solution** [For RPC Label Here] Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: RED 1/9/15 out of range during APM, performed solution change + rerun APM protocol.	R. Driscoll DMT TECH,
02/15/15	**DataMaster Simulator Solution** [For RP ce Label Here] Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: 02/15/2015 NJV Set clock to correct time	N. Vaillancourt
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011