

# SOLUTION CHANGE

DataMaster DMT: 122206  
Location: FRANKLIN COUNTY SO Enosburgh  
Date: 04/13/2016  
Performed by: NATHAN J VAILLANCOURT



## Accuracy and Precision Check

Concentration = 0.099 g/210L  
Lot # = 15-24-100  
Range = 0.094 - 0.104 *RES 4/14/16*  
Average = 0.101 g/210L  
Std Dev = 0.0013

Simulator Temperature: 34.0°C

Performed by Nathan Vaillancourt

Date 04/13/2016