

SOLUTION CHANGE

DataMaster DMT: 122406
Location: Essex PD
Date: 12/19/2015
Performed by: ROBERT J KISSINGER



Accuracy and Precision Check

Concentration = 0.097 g/210L
Lot # = 15-23-100
Range = 0.092 - 0.102
Average = 0.100 g/210L
Std Dev = 0.0000

ROD 12/21/15

Simulator Temperature: 33.9°C

Performed by



Date

12/19/2015