DATAMASTER DMT MAINTENANCE LOG

Wilmington PD
Department/Agency

300151
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/29/15	**DataMaster Simulator Solution** Lot # 14-12-100 EtOHaq Conc: 0.098 [For RPC Range: 0.093 - 0.103 Label Here] Preparation Date: 9/10/2014 Expires: 9/10/2015 OPENED: 7/29/15 RCD Installation post TSI	R. Driscoll DMT Tech.
10-23-15	**DataMaster Simulator Solution** Lot # 15-09-100 EtOHaq Conc: 0.099 g/210L [For RPC Range: 0.094 - 0.104 g/210L ce Label Here] Preparation Date: 6/24/2015 Expires: 6/24/2016 10-23-15 MAD OPENED: RPC - OCT - PASSED	M. DeMarco
	DataMaster Simulator Solution Lot # 15-09-100 EtOHaq Conc: 0.099 g/210L [FO Range: 0.094 - 0.104 g/210L e Place Label Here] Preparation Date: 6/24/2015 Expires: 6/24/2016 OPENED: 2/11/16 RPC - Feb 2016 PASSED	M. DeMarco
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011

Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
01/28/2015	08:34:02	WILMINGTON PD	MARK J DENAULT	0.000		0.095		
02/07/2015	04:32:00	WILMINGTON PD	PATRICK M BREWER	INCOMPLETE	0.105	0.095		RESPONSE TIMEOUT
02/14/2015	00:47:06	WILMINGTON PD	PATRICK M BREWER	REFUSED		0.095		
02/14/2015	06:37:17	WILMINGTON PD	PATRICK M BREWER	REFUSED		0.095		
02/28/2015	22:37:05	FISH & WILDLIFE	CODY N JACKMAN					Pump Error
02/28/2015	22:39:01	FISH & WILDLIFE	CODY N JACKMAN	0.087		0.094		
03/06/2015	17:09:46	WILMINGTON PD	REBECCA L MORRIS	0.067	INTERFERENCE	0.094	0.095	
03/15/2015	01:26:17	WILMINGTON PD	PATRICK M BREWER	0.194	0.187	0.094	0.094	
04/03/2015	13:33:55	wilmington	MARK J DENAULT	0.000		0.094		
04/29/2015	01:53:46	WILMINGTON PD	JASON M NOBLET	0.085	0.083	0.094	0.094	
06/14/2015	07:07:28	WILMINGTON PD	JOSEPH M SZAREJKO			0.090		Simulator Out of Range
07/02/2015	19:05:38	WILMINGTON PD	JASON M NOBLET	0.135		0.094		RESPONSE TIMEOUT
07/02/2015	19:15:14	WILMINGTON PD	JASON M NOBLET	0.136		0.095		
07/15/2015	20:30:10	WILMINGTON PD	JASON M NOBLET	0.154		0.095		
08/21/2015	23:43:37	WILMINGTON PD	JASON M NOBLET			0.097		RESPONSE TIMEOUT
08/21/2015	23:49:04	WILMINGTON PD	JASON M NOBLET	REFUSED		0.098		
09/03/2015	21:25:53	WILMINGTON PD	JASON M NOBLET	0.252		0.096		
09/09/2015	20:17:59	WILMINGTON PD	M MURANO	REFUSED		0.097		
09/13/2015	10:38:29	WILMINGTON PD	ADAM S BELVILLE	0.051		0.097		
09/17/2015	02:31:28	WILMINGTON PD	ADAM S BELVILLE	REFUSED		0.096		
09/27/2015	22:32:27	SHAFTSBURY VSP	JUSTIN W WALKER	0.082		0.096		
10/04/2015	11:11:51	WILMINGTON PD	JASON M NOBLET	0.031		0.095		
10/10/2015	18:54:04	WILMINGTON PD	JASON M NOBLET	0.172		0.097		
10/11/2015	19:01:23	WILMINGTON PD	M MURANO	0.196		0.095		
10/12/2015	15:39:03	WILMINGTON PD	JASON M NOBLET	0.240		0.097		
10/14/2015	00:25:29	WILMINGTON PD	M MURANO	REFUSED		0.095		
10/18/2015	22:43:46	BRATTLEBORO VSP	JOHN M WAITEKUS	0.254		0.095		
10/21/2015	19:07:55	WILMINGTON PD	CRAIG L WATROUS	0.216		0.094		
12/20/2015	23:23:01	WILMINGTON PD	M MURANO	0.122		0.096		
12/26/2015	19:48:53	BENNINGTON PD	KEITH R DIOTTE	0.152	0.152	0.095	0.095	
12/28/2015	04:15:07	WILMINGTON PD	M MURANO	0.162		0.095		
01/09/2016	10:26:44	WILMINGTON PD	LINDA A SHEDD	0.014		0.094		
01/28/2016	00:59:03	Wilmington Police	LINDA A SHEDD	REFUSED		0.095		
01/29/2016	10:50:49	WILMINGTON PD	MARK J DENAULT			0.093		Simulator Out of Range
01/29/2016	10:54:29	WILMINGTON PD	MARK J DENAULT	0.252		0.094		RESPONSE TIMEOUT

Status Record History

Status Date	Status Time	Test Type	Status Message
11/17/2011	10:17:58	DUI Subject Record	RESPONSE TIMEOUT
11/22/2011	08:35:27	Check-In Confirmation Record	Simulator Out of Range
12/12/2011	14:24:46	Accuracy & Precision Records	Blank Error
12/13/2011	08:13:22	Accuracy & Precision Records	Simulator Timed Out
12/13/2011	08:16:25	Accuracy & Precision Records	Simulator Timed Out
12/13/2011	08:20:41	Accuracy & Precision Records	Simulator Timed Out
12/13/2011	09:28:34	Accuracy & Precision Records	Simulator Timed Out
12/13/2011	09:31:36	Accuracy & Precision Records	Simulator Timed Out
12/13/2011	13:50:38	Accuracy & Precision Records	Simulator Timed Out
02/03/2013	02:11:50	DUI Subject Record	Simulator Out of Range
03/15/2013	18:54:38	DUI Subject Record	Simulator Out of Range
04/06/2013	19:29:27	DUI Subject Record	Simulator Out of Range
04/07/2013	09:34:34	Solution Change Record	ACCURACY AND PRECISION CHECK FAILED
09/22/2013	01:13:27	DUI Subject Record	RESPONSE TIMEOUT
09/24/2013	15:27:15	DUI Subject Record	Simulator Out of Range
12/31/2013	23:02:33	DUI Subject Record	Simulator Out of Range
12/31/2013	23:05:50	DUI Subject Record	Simulator Out of Range
01/30/2014	16:09:08	RPC Record	SAMPLE ACCEPTANCE TEST FAILED
02/21/2014	23:22:28	DUI Subject Record	Simulator Out of Range
11/25/2014	04:14:33	DUI Subject Record	RESPONSE TIMEOUT
12/14/2014	16:20:15	DUI Subject Record	RESPONSE TIMEOUT
01/01/2015	02:33:51	DUI Subject Record	Simulator Out of Range
01/23/2015	19:56:34	DUI Subject Record	Simulator Out of Range
02/07/2015	04:39:55	DUI Subject Record	RESPONSE TIMEOUT
02/28/2015	22:37:28	DUI Subject Record	Pump Error
06/14/2015	07:08:56	DUI Subject Record	Simulator Out of Range
07/02/2015	19:10:21	DUI Subject Record	RESPONSE TIMEOUT
08/21/2015	23:46:48	DUI Subject Record	RESPONSE TIMEOUT
01/29/2016	10:52:17	DUI Subject Record	Simulator Out of Range
01/29/2016	10:59:50	DUI Subject Record	RESPONSE TIMEOUT