

SOLUTION CHANGE

DataMaster DMT: 300387
Location: Hinesburg PD
Date: 06/26/2014
Performed by: BRIAN K FOX



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 13-21-100
Range = 0.094 - 0.104
Average = 0.098 g/210L
Std Dev = 0.0000

Simulator Temperature: 34°C

Performed by

B K F

Date

06/26/2014