VERMONT FORENSIC LABORATORY

103 South Main Street
PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
http://vfl.vermont.gov

Certificate of Preparation of DataMaster Interference Solution

On 2/4/15, an interference solution was prepared by Amanda Bolduc for use as a DataMaster Interference Check solution. The solution was prepared per standard laboratory practice and assigned the lot number 15-04-08A1 and an expiration date of 1/21/16.

This solution was prepared by adding 100 µL of acetone to 1 L of ethanol solution 15-04-080, which was certified on 1/29/15.

This solution is certified as presenting 0.080 g/210L simulated breath ethanol 0.01% vol/vol Acetone concentration and approved for use as a DataMaster Interference Check solution.

Dated at Waterbury, Vermont this 4th day of February, 2015.

A Bolduc

Amanda Bolduc

Reviewed By: [illegible] Date: 2/4/2015