VERMONT FORENSIC LABORATORY

103 South Main Street
PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
http://vfl.vermont.gov

Certificate of Preparation of DataMaster Interference Solution

On 09/29/2015, an interference solution was prepared by Jeffery Dukette for use as a DataMaster Interference Check solution. The solution was prepared per standard laboratory practice and assigned the lot number 15-16-04M and an expiration date of 09/29/2016.

This solution was prepared by adding 200 µL of methanol to 500 mL of diH_2O.

This solution is certified as presenting a 0.04% v/v methanol concentration and approved for use as a DataMaster Interference Check solution.

Dated at Waterbury, Vermont this 29th day of September, 2015.

Jeffery Dukette

Reviewed By: _____ Date: 9/30/2015