

VERMONT FORENSIC LABORATORY

103 South Main Street
PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
<http://vfl.vermont.gov>




Certificate of Preparation of DataMaster Interference Solution

On 10/14/2015, an interference solution was prepared by Jeffery Dukette for use as a DataMaster Interference Check solution. The solution was prepared per standard laboratory practice and assigned the lot number 15-18-040A1 and an expiration date of 09/29/2016.

This solution was prepared by adding 50 μ L of acetone to 500 mL of ethanol solution 15-17-040, which was certified on 10/6/2015.

This solution is certified as presenting 0.039 g/210L simulated breath ethanol and 0.01 % vol/vol acetone concentration and approved for use as a DataMaster Interference Check solution.

Dated at Waterbury, Vermont this 15th day of October, 2015.



Jeffery Dukette

Reviewed By:  Date: 10/15/15