

VERMONT FORENSIC LABORATORY

103 South Main Street
PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
<http://vfl.vermont.gov>




Certificate of Preparation of DataMaster Interference Solution

On 07/18/2016, an interference solution was prepared by Jeffery Dukette for use as a DataMaster Interference Check solution. The solution was prepared per standard laboratory practice and assigned the lot number 16-14-080A1 and an expiration date of 10/26/2016.

This solution was prepared by adding 100 µl of acetone to 1 L of ethanol solution 15-22-080, which was certified on 11/03/2015.

This solution is certified as presenting 0.079 g/210L simulated breath ethanol and 0.01% (vol/vol) acetone concentration and approved for use as a DataMaster Interference solution.

Dated at Waterbury, Vermont this 19th day of July, 2016.



Jeffery Dukette

Reviewed By:  _____ Date: 7/19/16