OHIO DEPARTMENT OF HEALTH

ALCOHOL AND DRUG TESTING PROGRAM

APPROVAL OF INSTRUMENT CHECK SOLUTION, OAC 370453-04

Lot or Batch # 12220 Target Value 0.101 g/210L

This instrument check solution is to be used in a simulator for the instrument checkof approved evidential breath testing instruments, as defined in OAC 370453-02.

This instrument check solution contains**1.234 mg/mL** ethyl alcohol in purified water. When used according to instrument check forms, it will produce a result at or within ±0005 g/210L of the target value in the approved breath testing instrument, when that instrument is in proper working condition.

This instrument check solution shall not be used more than three months after its date of firstuse or more than one year afterits date of manufacture, **November 20, 2013.** Whichever comes first.

This solution expires November 20, 2013

Approved this 3rd day of January, 2013

Director of Health