Digital Thermometer Performance Check

Reference Item	Reference Temperature	Test Temperature	Difference	Pass/Fail*
NIST Traceable QC Thermometer	33.9	34.1	0.2	Pass

* +/- 0.3 degree Celsius acceptable range

Certified Thermometer Serial Number 91A-4704

Thermometer Serial Number 111814517

Date 12/19/2014

Test Perfomed by Rebecca Mead RCM

Reviewer 12/22/2014

VFL Thermometer blank template: public/lab1/QC/QAQCfolder/thermometers