Digital Thermometer Performance Check

Item	Reference Temperature	Test Temperature	Difference	Pass/Fail*
NIST Traceable QC Thermometer	33.7	33.9	0.2	Pass

* +/- 0.3 degree Celsius acceptable range

Certified Reference Thermometer Serial Number: 91A-4704

Test Thermometer Serial Number: 140722257

Date: 11/6/2015

Test Performed By: Jeff Dukette

Reviewer: 11/6/15

Template path: Public\Lab1\Sectionfolders\QA\QA\QAQCfolder\Thermometers